



CALIBRATION CERTIFICATE

Customer name : RION SERVICE CENTER CO., LTD.
Customer address : 2-22-2, Hyoe, Hachioji, Tokyo 192-0918, JAPAN
Product : SOUND CALIBRATOR
Type : NC-75
Serial number : 00000000
Manufacturer : RION CO., LTD.
Calibration quantities : Sound pressure level (with reference standard microphone)
Calibration method : Measured by specified secondary standard microphone
according to JCSS calibration procedure
specified by RION SERVICE CENTER.
Ambient conditions : Temperature 23.0 °C, Humidity 50 %R.H.,
Static pressure 101.33 kPa
Calibration date : **/**/**** (DD/MM/YYYY)
Calibration location : 2-22-2, Hyoe, Hachioji, Tokyo 192-0918, JAPAN
Calibration Room

We hereby certify that the results of this calibration were as follows.

Issue date : **/**/**** (DD/MM/YYYY)

○○○○○ ○○○○○ Head of Office

Quality Control Office
Engineering Business Unit
RION SERVICE CENTER CO., LTD.
2-22-2, Hyoe, Hachioji, Tokyo 192-0918, JAPAN

This certificate is based on article 144 of the Measurement Law and indicates the result of calibration in accordance with measurement standards traceable to Primary Measurement Standards (National Standards) which realizes the physical units of measurement according to the International System of Units (SI).

The accreditation symbol is attestation of which the result of calibration is traceable to Primary Measurement Standards (National Standards).

The certificate shall not be reproduced except in full, without the written approval of the issuing laboratory.

The calibration laboratory who issued this calibration certificate conforms to ISO/IEC 17025:2017.

This calibration certificate was issued by the calibration laboratory accredited by IAJapan who is a signatory to the Mutual Recognition Arrangement (MRA) of International Laboratory Accreditation Cooperation (ILAC) and Asia Pacific Accreditation Cooperation (APAC). This (These) calibration result(s) may be accepted internationally through ILAC/APAC MRA.

CALIBRATION RESULT

1. Sound pressure level (with reference standard microphone)

Measured value	Expanded uncertainty*1
94.00 dB	0.09 dB

Specified secondary standard microphone :

Type : BK4160

Serial number : 0000000

Reference Sound pressure : 2×10^{-5} Pa

*1 Defines an interval estimated to have a level of confidence of approximately 95%

Coverage factor $k=2$

Calibration result is the calibration value in ambient conditions during calibration.

BE OUT OF JCSS CALIBRATION

1. Frequency

Measured value	Measurement uncertainty ($k=2$)
1000.0 Hz	5×10^{-7} Hz

Working measurement standard universal counter :

Type : 53132A

Serial number : 0000000000

(A2LA Calibration Certificate No.000-00000)

2. Total distortion

Measured value
0.1 %

Working measurement standard distortion meter :

Type : AD725D

Serial number : 0000000000

(A2LA Calibration Certificate No.000-00000)

End of certificate