

CALIBRATION CERTIFICATE

Customer Name: *****
Description: SOUND CALIBRATOR
Model Name: NC-75
Serial Number: *****
Calibration Date: 06/06/2023 (DD/MM/YYYY)
Ambient condition: Temperature 23 °C Relative Humidity 50 %RH

We hereby certify that the above product was tested and calibrated according to the prescribed RION SERVICE CENTER (RSC) procedures, and that it fulfills all requirements of the product specifications, as described in the attached test report. The measuring equipment and reference devices used for testing and calibrating this unit are managed under the RSC traceability system and are traceable according to official Japanese standards and official standards of countries belonging to the International Committee of Weights and Measures.

PRIMARY STANDARDS

Model Description	Model Number	Control Number	Cal Due Date MM/YYYY
Condenser Microphone	BK4160	2545030	12/2023
Universal Counter	53132A	MY40007353	10/2023
Automatic Distortion Analyzer	AD725D	3000032468	12/2023
Digital multimeter	3458A	MY45051584	07/2023

WORK STANDARDS

Model Description	Model Number	Control Number	Cal Due Date MM/YYYY
Measuring Amplifier	NA-42改	RH-815	08/2023

RION SERVICE CENTER CO., LTD.
2-22-2 Hyoe, Hachioji, Tokyo 192-0918, JAPAN

Head of Office, Quality Control Office _____